
40 C.F.R. § 61.245

Test methods and procedures.

- (a) Each owner or operator subject to the provisions of this subpart shall comply with the test methods and procedures requirements provided in this section.
- (b) Monitoring, as required in §§ 61.242, 61.243, 61.244, and 61.135, shall comply with the following requirements:
 - (1) Monitoring shall comply with Method 21 of appendix A of 40 CFR part 60.
 - (2) The detection instrument shall meet the performance criteria of Method 21.
 - (3) The instrument shall be calibrated before use on each day of its use by the procedures specified in Method 21.
 - (4) Calibration gases shall be:
 - (i) Zero air (less than 10 ppm of hydrocarbon in air); and
 - (ii) A mixture of methane or n-hexane and air at a concentration of approximately, but less than, 10,000 ppm methane or n-hexane.
 - (5) The instrument probe shall be traversed around all potential leak interfaces as close to the interface as possible as described in Method 21.
- (c) When equipment is tested for compliance with or monitored for no detectable emissions, the owner or operator shall comply with the following requirements:
 - (1) The requirements of paragraphs (b) (1) through (4) shall apply.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)