

40 C.F.R. § 1065.295

PM inertial balance for field-testing analysis.

- (a) *Application.* You may use an inertial balance to quantify net PM on a sample medium for field testing.
- (b) *Component requirements.* We recommend that you use a balance that meets the specifications in Table 1 of § 1065.205. Note that your balance-based system must meet the linearity verification in § 1065.307. If the balance uses an internal calibration process for routine spanning and linearity verifications, the process must be NIST-traceable.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)