

40 C.F.R. § 160.63

Maintenance and calibration of equipment.

- (a) Equipment shall be adequately inspected, cleaned, and maintained. Equipment used for the generation, measurement, or assessment of data shall be adequately tested, calibrated, and/or standardized.
- (b) The written standard operating procedures required under § 160.81(b)(11) shall set forth in sufficient detail the methods, materials, and schedules to be used in the routine inspection, cleaning, maintenance, testing, calibration, and/ or standardization of equipment, and shall specify, when appropriate, remedial action to be taken in the event of failure or malfunction of equipment. The written standard operating procedures shall designate the person responsible for the performance of each operation.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)