
40 C.F.R. § 86.116–94

Calibrations, frequency and overview.

- (a) Calibrations shall be performed as specified in § 86.117 through § 86.126.
- (b) At least yearly or after any maintenance which could alter background emission levels, evaporative enclosure background emission measurements shall be performed.
- (c) At least monthly or after any maintenance which could alter calibration, the following calibrations and checks shall be performed:
 - (1) Calibrate the THC analyzers (both evaporative and exhaust instruments), methane analyzer, carbon dioxide analyzer, carbon monoxide analyzer, and oxides of nitrogen analyzer (certain analyzers may require more frequent calibration depending on particular equipment and uses).

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)