

---

# 40 C.F.R. § 86.514-78

---

## Analytical gases.

---

- (a) *Analyzer gases.* (1) Gases for the CO and CO<sub>2</sub> analyzers shall be single blends of CO and CO<sub>2</sub> respectively using nitrogen as the diluent.
- (2) Gases for the THC analyzer shall be:
- (i) Single blends of propane using air as the diluent; and
  - (ii) Optionally, for response factor determination, single blends of methanol using air as the diluent.
- (3) Gases for the NO<sub>x</sub> analyzer shall be single blends of NO named as NO<sub>x</sub> with a maximum NO<sub>2</sub> concentration of 5 percent of the nominal value using nitrogen as the diluent.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)