
40 C.F.R. § 63.363

Compliance and performance provisions.

- (a) *Continuous compliance.* You must demonstrate continuous compliance with the applicable emission standard(s) using an EtO CEMS, including a shared EtO CEMS, installed and operated in accordance with the requirements of Performance Specification 19 in appendix B and Procedure 7 in appendix F to part 60 of this chapter. Alternatively, if you own or operate a facility where EtO use is less than 100 pounds/yr, you may demonstrate continuous compliance by conducting annual performance tests using the performance testing requirements in § 63.7, according to the applicability in table 6 to this subpart, the procedures listed in this section, and the test methods listed in § 63.365. If you elect to demonstrate compliance through periodic performance testing, you must also demonstrate continuous compliance with each operating limit required under this section according to the methods specified in § 63.364. If you own or operate an area source facility where EtO use is less than 100 pounds/yr where an existing collection of Group 2 room air emission is operated in accordance with the PTE requirements of EPA Method 204 of appendix M to part 51 of this chapter, you may instead conduct these performance tests once every three years.
- (b) *Initial compliance for Facilities that use EtO CEMS.* To demonstrate initial compliance with an emission standard using a CEMS that measures HAP concentrations directly (*i.e.*, an EtO CEMS), the initial performance test must consist of the first 30 operating days after the certification of the CEMS according to Performance Specification 19 in Appendix B to part 40 of this chapter. The initial compliance demonstration period must be completed on or before the date that compliance must be demonstrated (*i.e.*, 180 days after the applicable compliance date). You must follow the procedures in appendix A to this subpart.

This document is only available to subscribers. Please log in or purchase access.

[Purchase Login](#)