

42 C.F.R. § 84.144

Hand-operated blower test; minimum requirements.

- (a) Hand-operated blowers shall be tested by attaching them to a mechanical drive and operating them 6 to 8 hours daily for a period of 100 hours at a speed necessary to deliver 50 liters of air per minute through each completely assembled respirator. Each respirator shall be equipped with the maximum length of hose with which the device is to be approved and the hose shall be connected to each blower or manifold outlet designed for hose connections.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)