

42 C.F.R. § 84.145

Motor-operated blower test; minimum requirements.

- (a) Motor-operated blowers shall be tested by operating them at their specified running speed 6 to 8 hours daily for a period of 100 hours when assembled with the kind and maximum length of hose for which the device is to be approved and when connected to each blower or manifold outlet designed for hose connections.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)