

## 42 C.F.R. § 84.155

---

### Airflow resistance test; Type C supplied-air respirator, continuous flow class and Type CE supplied-air respirator; minimum requirements.

---

The resistance to air flowing from the respirator shall not exceed 25 mm. (1 inch) of water-column height when the air flow into the respiratory-inlet covering is 115 liters (4 cubic feet) per minute.

This document is only available to subscribers. Please log in or purchase access.

[Purchase Login](#)