

42 C.F.R. § 84.156

Airflow resistance test; Type C supplied-air respirator, demand class; minimum requirements.

- (a) Inhalation resistance shall not exceed 50 millimeters (2 inches) of water at an air flow of 115 liters (4 cubic feet) per minute.

This document is only available to subscribers. Please log in or purchase access.

[Purchase Login](#)