

42 C.F.R. § 84.207

Bench tests; gas and vapor tests; minimum requirements; general.

- (a) Bench tests will be made on an apparatus that allows the test atmosphere at 50 ± 5 percent relative humidity and room temperature, approximately 25°C , to enter the cartridges continuously at predetermined concentrations and rates of flow, and that has means for determining the test life of the cartridges.
- (b) Where two cartridges are used in parallel on a chemical cartridge respirator, the bench test will be performed with the cartridges arranged in parallel, and the test requirements will apply to the combination rather than to the individual cartridges.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)