

42 C.F.R. § 84.84

Hand-operated valves; minimum requirements.

- (a) Hand-operated valves shall be designed and constructed to prevent removal of the stem from the valve body during normal usage to insure against a sudden release of the full pressure of the container when the valve is opened.
- (b) Valves shall be designed or positioned to prevent accidental opening and closing, and damage from external forces.
- (c) Valves operated during use of the apparatus shall be installed in locations where they can be readily adjusted by the wearer.

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)