

42 C.F.R. § 84.94

Gas flow test; closed-circuit apparatus.

- (a) Where oxygen is supplied by a constant-flow device only, the rate of flow shall be at least 3 liters per minute for the entire rated service time of the apparatus.

This document is only available to subscribers. Please log in or purchase access.

[Purchase Login](#)