

## 42 C.F.R. § 486.110

---

### Condition for coverage: Inspection of equipment.

---

Inspections of all X-ray equipment and shielding are made by qualified individuals at intervals not greater than every 24 months.

- (a) *Standard—qualified inspectors.* Inspections are made at least every 24 months by a radiation health specialist who is on the staff of or approved by an appropriate State or local government agency.

This document is only available to subscribers. Please log in or purchase access.

[Purchase Login](#)