

## 42 C.F.R. § 493.837

---

### Standard; General immunology.

---

- (a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.
- (b) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.
- (c) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if—

This document is only available to subscribers. Please [log in](#) or [purchase access](#).

[Purchase Login](#)